

Trusted since 1917

## **MICROBIOLOGY MEMBRANE FILTER TECHNIQUE**

### **ADVANTEC® MCE Membrane Filters**

Advantec MCE membranes are manufactured to ensure superior guality and reproducibility. All 0.45µm white gridded membranes are tested for Coliform, Fecal Streptococci, and Serratia marcescens. All 0.65µm are tested to ensure complete retention and optimal recovery of Fecal Coliform and Saccharomyces cerevisiae. Our Black and Green membranes are tested for optimal recovery of yeast and total bacteria. Additional tests are performed for uniform wetting, freedom from grid line inhibition, and optimal color reactions on appropriate test media. Quality Assurance Certification is available for individual lot numbers...

### **Application:**


Ideal for capturing and culturing microorganisms using the MF Technique in the Water and Wastewater Industry.

TEST ORGANISM	TEST METHOD	PORE/RETENTION
Total Coliforms	9222 B	0.45 μm
Fecal Coliforms	9222 D	0.45 μm
	9230 C	
Total Bacteria		0.20 μm
	9215 D	0.45 μm
Salmonella	9260 B	0.45 μm
Pseudomonas sp	9213	0.20 μm
(Pseudomonas	9213 E	0.80 µm
aeruginosa)		

### Add Media Place filt he pad



#### Sample Filtration and Growth Procedures



SUPPORTING PRODUCTS:			
<b>Petri-dishes</b> With Pad Without Pad	<b>Item (Cat. No.)</b> PD-47B (42004020) PD-47A (42004010)		
Absorbent Pads	B200G047A (02511049)		
Filterceps	FS-1 (17800000)		
SS Manifold w/ 3-Way Valve			
3-Place 6-Place	KM-3N (17313100) KM-6N (17313300)		
47mm Holder w/ Funne Glass base, glass frit Glass base, ss screen Stainless Steel Polysulfone			



### **Specifications & Ordering Information**

47mm Sterile Membranes, Gridded (100/pk)

		ltem		
Pore size	Color	Without Pad	With Pad	
0.20 μm	white	A020H047A	A020F047A	
0.45 µm	white	A045H047A	A045F047A	
	white	A045H047A-P	-	
	black	A045R047A	A045M047A	
		A045R047A-P	-	
	green	A045W047A	-	
0.65 µm	white	A065H047A	A065F047A	
0.80 µm	white	A080H047A	A080F047A	
	black	A080R047A	A080M047A	

#### Other grades, diameters and pack quantities are available.





#### Sample Growth Results



A045W047A Lactic acid bacteria cultured in growth media A045F047A Fecal E. coli cultured in m-FC media

A080R047A

Saccharomyces cultured

in m-Green Yeast and Mold Broth



## For Water Quality

#### Trusted since 1917

### TSS TESTING -- Standard Methods 2540

### ADVANTEC<sup>®</sup> Glass Fiber Filters: Grade GA-55

Advantec GA-55 Glass Fiber Filters are free of binders and made of pure, fine borosilicate glass. These high purity filters have a fast flow rate and a high loading capacity.

### Application

Ideal for TSS (Total Suspended Solid) Analysis in the Water and Wastewater Industry.

### **Specifications**

	low Time <sup>1</sup>		Temp.
0.6 μm 55 g/m <sup>2</sup> 0.21 mm	23 sec.	None	500°C

<sup>1</sup>Flow time is the time in seconds to filter 1,000 mL of distilled water at 20°C under differential pressure of 39 kPa through a 9.6 cm<sup>2</sup> section of filter.

Ordering Information					
Catalog No.	ltem	Catalog No.	ltem	Catalog No.	ltem
36261024	GA-55 24mm dia.	36261047	GA-55 47 mm dia.	36261090	GA-55 90 mm dia.
36261025	GA-55 25 mm dia.	36261055	GA-55 55 mm dia.	36261110	GA-55 110 mm dia.

## OIL & GREASE TESTING -- Standard Methods 5520

### **ADVANTEC**<sup>®</sup> Quantitative Filter Papers: Grade No.5B

Advantec Quantitative Filter Papers are made of 99% alpha cotton cellulose which has a high degree of polymerization. The low 0.01% ash content makes it the preferred filter for quantitative analysis.

### Application

Ideal for Oil and Grease testing in the Water and Wastewater Industry.



Nominal Rating	Thickness	Weight	Flow Time <sup>1</sup>	Absorption Speed <sup>2</sup>	Wet Strength <sup>3</sup>
4 μm (Medium)	0.21 mm	108 g/m <sup>2</sup>	195 sec.	7.0 cm	12 kPa

<sup>1</sup>Flow time is the time in seconds required to filter 100 mL of distilled water at 20°C under pressure supplied by a 10 cm water column through a 10 cm<sup>2</sup> section of filter paper. <sup>2</sup>Absorption speed is the distance in cm that water will travel in an upright strip of filter paper in 10 minutes at 20°C. <sup>3</sup>Wet Strength is the pressure measured by Mullen Burst Strength Tester after soaking in water.

### **Ordering Information**

Catalog No.	Item	Catalog No.	Item	
01521110	No.5B 110 mm dia.	01521150	No.5B 150 mm dia.	
01521125	No.5B 125 mm dia.	01521185	No.5B 185 mm dia.	
Additional grades and diameters are available. Please contact us				

Additional grades and diameters are available. Please contact us.

Standard Methods for the examination of water and wastewater is the publication by the American Public Health Association (APHA), the American Water Works Association (AWWA) and the Water Environment Federation (WEF).

Contact:

# **ADVANTEC**<sup>®</sup>

#### ADVANTEC TOYO KAISHA, LTD.

Hibiya-Kokusai BLDG, 2-2-3, Uchisaiwaicho, Chiyoda-ku, Tokyo, 100-0011 Japan Phone +81-3-5521-2181 Facsimile +81-3-5521-2182 Email atk-otd@ADVANTEC.co.jp Website www.ADVANTEC.co.jp/en/



FILTER PAPER

DVANTEC

<sup>100</sup> CIRCLES **150**mm